



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY, BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS, THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 1 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Current at 50 Hz	Using Multiproduct Calibrator by Direct method	1 A to 5 A	0.0015 A to 0.02 A
2	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Current at 50 Hz	Using Multiproduct Calibrator by Direct method	100 mA to 1 A	0.2 mA to 0.0015 A
3	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Current at 50Hz	Using Multiproduct Calibrator by Direct method	33 mA to 100 mA	0.04 mA to 0.2 mA
4	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage at 50 Hz	Using Multiproduct Calibrator by Direct method	100 V to 600 V	0.09 V to 0.43 V
5	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage at 50 Hz	Using Multiproduct Calibrator by Direct method	330 mV to 4 V	0.21 mV to 0.007 V



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY,
BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS,
THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 2 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
6	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage at 50 Hz	Using Multiproduct Calibrator by Direct method	4 V to 100 V	0.007 V to 0.09 V
7	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	Using Digital Multimeter by Direct method	1 A to 5 A	1.5 mA to 12.9 mA
8	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	Using Digital Multimeter by Direct method	10 mA to 100 mA	13.6 µA to 113.2 µA
9	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	Using Digital Multimeter by Direct method	100 mA to 1 A	0.12 mA to 1.5 mA
10	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Resistance	Using Digital Multimeter by Direct method	1 kohm to 100 kohm	0.51 Ohm to 50.6 Ohm
11	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Resistance	Using Digital Multimeter by Direct method	10 Ohm to 100 Ohm	0.025 Ohm to 0.05 Ohm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY,
BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS,
THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 3 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
12	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Resistance	Using Digital Multimeter by Direct method	100 kohm to 1 Mohm	50.6 Ohm to 0.69 kohm
13	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Resistance	Using Digital Multimeter by Direct method	100 Ohm to 400 Ohm	0.05 Ohm to 0.44 Ohm
14	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Resistance	Using Digital Multimeter by Direct method	400 Ohm to 1 kohm	0.44 Ohm to 0.51 Ohm
15	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using Digital Multimeter by Direct method	20 mV to 400 mV	0.013 mV to 0.16 mV
16	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using Digital Multimeter by Direct method	4 V to 40 V	1.7 mV to 15 mV
17	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using Digital Multimeter by Direct method	40 V to 400 V	15 mV to 105 mV



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY,
BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS,
THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2574

Page No

4 of 17

Validity

14/02/2023 to 13/02/2025

Last Amended on

10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
18	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using Digital Multimeter by Direct method	400 mV to 4 V	0.16 mV to 1.7 mV
19	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multiproduct Calibrator by Direct method	1 A to 10 A	0.001 A to 0.01 A
20	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multiproduct Calibrator by Direct method	10 mA to 100 mA	0.002 mA to 0.031 mA
21	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using Multiproduct Calibrator by Direct method	100 mA to 1 A	0.031 mA to 0.001 A
22	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using Multiproduct Calibrator by Direct method	1 kohm to 100 kohm	0.13 Ohm to 13 Ohm
23	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using Multiproduct Calibrator by Direct method	10 Ohm to 400 Ohm	0.013 Ohm to 0.07 Ohm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY,
BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS,
THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 5 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
24	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using Multiproduct Calibrator by Direct method	100 kohm to 1 Mohm	13 Ohm to 190 Ohm
25	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using Multiproduct Calibrator by Direct method	400 Ohm to 1 kohm	0.07 Ohm to 0.13 Ohm
26	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using Multiproduct Calibrator by Direct method	20 mV to 400 mV	0.007 mV to 0.04 mV
27	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using Multiproduct Calibrator by Direct method	4 V to 40 V	0.26 mV to 3.3 mV
28	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using Multiproduct Calibrator by Direct method	40 V to 400 V	3.3 mV to 25 mV
29	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using Multiproduct Calibrator by Direct method	400 mV to 4 V	0.04 mV to 0.26 mV



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY,
BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS,
THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 6 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
30	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Digital Multimeter by Direct method	1 kHz to 5 kHz	0.22 Hz to 1.5 Hz
31	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Digital Multimeter by Direct method	5 kHz to 50 kHz	1.5 Hz to 14 Hz
32	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Digital Multimeter by Direct method	50 Hz to 1 kHz	0.32 Hz to 0.22 Hz
33	ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)	Frequency	Using Multiproduct Calibrator by Direct method	5 kHz to 50 kHz	0.1 Hz to 5.9 Hz
34	ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)	Frequency	Using Multiproduct Calibrator by Direct method	50 Hz to 5 kHz	0.013 Hz to 0.1 Hz
35	ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)	Frequency	Using Multiproduct Calibrator by Direct method	50 kHz to 100 kHz	5.9 Hz to 58.6 Hz



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY,
BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS,
THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 7 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
36	MECHANICAL-VOLUME	Micropipettes	Using weighing balance of readability: 0.01 mg and distilled water by Gravimetric method based on ISO 8655	>1 ml to 5 ml	2 μ l
37	MECHANICAL-VOLUME	Micropipettes	Using weighing balance of readability: 0.01 mg and distilled water by Gravimetric method based on ISO 8655	>100 μ l to 200 μ l	0.22 μ l
38	MECHANICAL-VOLUME	Micropipettes	Using weighing balance of readability: 0.01 mg and distilled water by Gravimetric method based on ISO 8655	>200 μ l to 1000 μ l	1.44 μ l
39	MECHANICAL-VOLUME	Micropipettes	Using weighing balance of readability: 0.01 mg and distilled water by Gravimetric method based on ISO 8655	10 μ l to 100 μ l	0.22 μ l



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY, BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS, THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 8 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
40	MECHANICAL-VOLUME	Volume Glassware, Burette, Pipettes, Measuring Cylinder, Standard Flasks	Using weighing balance of readability: 0.1 mg and distilled water as per ISO 4787: 2021	>100 ml to 500 ml	0.20ml
41	MECHANICAL-VOLUME	Volume Glassware, Burette, Pipettes, Measuring Cylinder, Standard Flasks, venous blood, hypodermic syringes (for non medical purpose)	Using weighing balance of readability: 0.1 mg and distilled water as per ISO 4787: 2021	>10 ml to 20 ml	15µl
42	MECHANICAL-VOLUME	Volume Glassware, Burette, Pipettes, Measuring Cylinder, Standard Flasks, venous blood, hypodermic syringes (for non medical purpose)	Using weighing balance of readability: 0.1 mg and distilled water as per ISO 4787: 2021	>20 ml to 100 ml	21µl
43	MECHANICAL-VOLUME	Volume Glassware, Burette, Pipettes, Measuring Cylinder, Standard Flasks, venous blood, hypodermic syringes (for non medical purpose)	Using weighing balance of readability: 0.1 mg and distilled water as per ISO 4787: 2021	1 ml to 10 ml	5µl



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY,
BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS,
THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 9 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
44	MECHANICAL-WEIGHTS	Weights: M2 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01g by ABA method as per OIML R-111	500 g	13.5mg
45	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	1 g	0.07mg
46	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01g by ABA method as per OIML R-111	1 kg	13.5mg
47	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	1 mg	0.06mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY, BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS, THIRUVANANTHAPURAM, KERALA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	CC-2574	Page No	10 of 17
Validity	14/02/2023 to 13/02/2025	Last Amended on	10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
48	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	10 g	0.07mg
49	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	10 mg	0.06mg
50	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	100 g	0.09mg
51	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	100 mg	0.06mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY, BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS, THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 11 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
52	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	2 g	0.07mg
53	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01g by ABA method as per OIML R-111	2 kg	14.5mg
54	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	2 mg	0.06mg
55	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	20 g	0.07mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY,
BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS,
THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 12 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
56	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	20 mg	0.06mg
57	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	200 g	0.12mg
58	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	200 mg	0.06mg
59	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	5 g	0.07mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY,
BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS,
THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 13 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
60	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	5 mg	0.06mg
61	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	50 g	0.08mg
62	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	50 mg	0.06mg
63	MECHANICAL-WEIGHTS	Weights:M1 Class weights and coarser	Using E2 class standard weights and weighing Balance of readability: 0.01 mg by ABA method as per OIML R-111	500 mg	0.06mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY, BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS, THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 14 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
64	THERMAL-SPECIFIC HEAT & HUMIDITY	RH meters and Thermo hygrometer	Using Thermo hygrometer and salt bath as per ASTM E104 by Comparison method	53 %rh	2%rh
65	THERMAL-SPECIFIC HEAT & HUMIDITY	RH meters and Thermo hygrometers	Using Thermo-hygrometer and salt bath as per ASTM E104 by Comparison method	45 %rh	2%rh
66	THERMAL-SPECIFIC HEAT & HUMIDITY	RH meters and Thermo hygrometers	Using Thermo hygrometer and salt bath as per ASTM E104 by Comparison method	75 %rh	2%rh
67	THERMAL-SPECIFIC HEAT & HUMIDITY	RH meters and Thermo hygrometers	Using Thermo-hygrometer and salt bath as per ASTM E104 by Comparison method	83 %rh	2%rh
68	THERMAL-TEMPERATURE	Liquid in Glass Thermometers	Using liquid bath and Secondary Precision Resistance Thermometers with Indicator by Comparison method	-20 °C to 200 °C	0.3°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY,
BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS,
THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2574

Page No

15 of 17

Validity

14/02/2023 to 13/02/2025

Last Amended on

10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
69	THERMAL-TEMPERATURE	Thermocouple, RTD, Thermistors With Indicator, Temperature Gauges	Using liquid bath, dry block Calibrator and Secondary Precision Resistance Thermometers with Indicator by Comparison method	-20 °C to 200 °C	0.3°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY, BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS, THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2574 **Page No** 16 of 17

Validity 14/02/2023 to 13/02/2025 **Last Amended on** 10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Site Facility					
1	MECHANICAL-WEIGHING SCALE AND BALANCE	Balance Class I & Coarser Readability 0.01 mg	Using E2 Class Standard Weights by Comparison method as per OIML R 76-1	1 mg to 1 g	0.06mg
2	MECHANICAL-WEIGHING SCALE AND BALANCE	Balance Class I & Coarser Readability 0.1 mg	Using E2 Class Standard Weights by Comparison method as per OIML R 76-1	10 mg to 500 g	1.07mg
3	MECHANICAL-WEIGHING SCALE AND BALANCE	Balance Class II & Coarser Readability 1 mg	Using E2 Class Standard Weights by Comparison method as per OIML R 76-1	200 mg to 500 g	1.07mg
4	MECHANICAL-WEIGHING SCALE AND BALANCE	Balance Class II & Coarser Readability: 10 mg	Using E2 Class Standard Weights by Comparison method as per OIML R 76-1	100 g to 2 kg	0.65g
5	THERMAL-TEMPERATURE	Liquid in Glass Thermometers	Using liquid bath and Secondary Precision Resistance Thermometers with Indicator by Comparison method	-20 °C to 200 °C	0.3°C
6	THERMAL-TEMPERATURE	Ovens,Baths	Using RTD with Data Acquisition system as per ASTM E 145 by Comparison method	-20 °C to 200 °C	2.0°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES & TECHNOLOGY,
BIOMEDICAL TECHNOLOGY WING, SATELMOND PALACE CAMPUS,
THIRUVANANTHAPURAM, KERALA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2574

Page No

17 of 17

Validity

14/02/2023 to 13/02/2025

Last Amended on

10/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
7	THERMAL-TEMPERATURE	Thermocouple, RTD, Thermistors With Indicator, Temperature Gauges	Using liquid bath, dry block Calibrator and Secondary Precision Resistance Thermometers with Indicator by Comparison method	-20 °C to 200 °C	0.3°C

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.