



श्रीचित्रातिरुनालआयुर्विज्ञानऔरप्रौद्योगिकीसंस्थान तिरुवनंतपुरम- 11, केरल
SreeChitraTirunal Institute for Medical Sciences and Technology, Thiruvananthapuram- 11, Kerala
(भारतसरकारकेअधीनराष्ट्रीयमहत्वसंस्थान)
(An Institute of National Importance under Government of India)
टेलीफॉननं./Telephone No. 0471-2443152 फाक्स/Fax 0471-24464332550728
ई-मेल/E-mail :sct@scimst.ac.inवेबसाइट/ Website : www.scimst.ac.in

CORRIGENDUM -2 dtd 08.02.2022

TENDER NO.SCT/H/IMP-IND/P1/2021-22/9

A. Date of Submission of Bid (Online and Offline) and Tender Opening date is changed.		
Particulars	Dates given as per corrigendum -1	Date and Time changed to
Last date and time of online submission of bids	15/02/2022 up to 5.00 PM	02/03/2022 up to 5.00 PM
Last date and time of submission of Hardcopy of Techno-commercial Bid with supporting documents (price bid has to be submitted online only). <i>The tender will stand rejected if the price bid is submitted along with hardcopy of techno-commercial bid</i>	19/02/2022 up to 01.00 PM	07/03/2022 up to 01.00 PM
Date of tender Opening	21/02/2022 at 2.30 PM	09/03/2022 at 2.30 PM

B. NEW MEMBERS - INDEPENDENT EXTERNAL MONITORS

Sl.No.	Independent External Monitors as per tender notice dtd.21.102021	New Members
Pg.No.2 of tender notice	Sri. Sharda Prasad, IPS (Rtd.)	Sri. Prahlad Kumar Sinha, IP & TAFS (Rtd.) Ph.No.09423677066 E-mail- pekay66@gmail.com
	Sri. Sanjeev Behari, IRS (Rtd.)	Sri. Javeed Ahmad, IPS (Rtd.) Ph.No.08527249595 E-mail - Javeed60@yahoo.com

C. PART-II of Annexure-1

Sl. No.	Description	To be read as
6.B	The tabletop shall have length of 300 cm, width - 20 cms at head area, 45 cms at trunk area and 65 cms at tail area, widest available table should be provided.	The widest available table should be provided. The tabletop shall have length of 280 cm, width - 20cms at head area, 45 cms at the trunk area and 52 cms at the tail area
12.U.12	(Sl.No.10) Computational flow dynamic (CFD) evaluation, (Sl.No.11) artificial intelligence powered applications for dose reduction, image analysis and diagnosis (Sl.No.12) Facility for robotic assisted intervention to be specified. Associated driving cartridge and integrating software to be quoted.	Sl No. 10 and 11, is optional Sl.No.12 (Already under the head of optional item)
12.V.11	All image manipulation and post processing should be possible from the examination room also. WiFi enabled mouse to be provided in the examination room for 3D image processing and evaluation.	Wi-Fi-enabled mouse or similar device should be provided in the examination room for 3D image processing and evaluation.
13.F	Provide a local storage system/local mini PACS system to store all the image data without any data loss. The storage size should be adequate enough to store all image data to be generated over 10 years. (Procedure load - 75 diagnostic cerebral angiogram and 25 neuro interventions per month)	Provide a local storage system/local mini PACS system to store all the image data without any data loss. The storage size should be adequate enough to store all image data to be generated over 10 years.(Procedure load - 75 diagnostic cerebral angiogram and 25 neuro interventions per month). A minimum of 20TB is required for storage.

D. Techno-commercial bid- Check List

27	Have you kept validity of 270 days from the Techno Commercial Bid Opening date as per the bidding document?	Have you kept validity of 180 days from the Techno Commercial Bid Opening date as per the bidding document?
----	---	---

E. Technical Compliance and check list is revised and is attached

Sd/-
Director