



**SREE CHITRA TIRUNAL INSTITUTE FOR MEDICAL SCIENCES &
TECHNOLOGY, TRIVANDRUM-695011, KERALA, INDIA**
(An Institute of National Importance under Government of India)
Phone:0471-2524445/145/225/425, Fax:0471-2550728 Email:purchase@sctimst.ac.in

CORRIGENDUM - TENDER NO. SCT/IMP/002/2018-19/SA16

INVERTED FLOURESCENCE MICROSCOPE

SPECIFICATION IN TENDER	TO BE READ AS
CAMERA: High performance microscopy camera with driver software, USB3.0 PCIE X1 interface, dual USB 3.0/USB 2.0 CABLE 3 M and BK7 protection glass (coated) 2.8 MEGA pixels or better.	CAMERA: High performance microscopy camera, dual USB 3.0/USB 2.0 CABLE 3 M and protection glass (coated) 2.8 MEGA pixels or better. Camera should be Peltier cooled CCD Monochrome camera with USB3.0 interface and protection glass coated.
MODULE FOR OPTICAL SECTIONING:	'Optical sectioning module should be factory inbuilt with the microscope and should be provided by same manufacturer.' - This clause is deleted.

Last date of receipt of tenders: 29.09.2018 upto 1.00pm

The tenders will be opened at 2.30PM on 01.10.2018.