



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2023/B/3547972
Dated/दिनांक : 20-06-2023

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	30-06-2023 16:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	30-06-2023 16:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Science And Technology
Department Name/विभाग का नाम	Department Of Science And Technology (dst)
Organisation Name/संगठन का नाम	Sree Chitra Tirunal Institute For Medical Sciences And Technology (sctimst)
Office Name/कार्यालय का नाम	Thiruvananthapuram
Total Quantity/कुल मात्रा	4250
Item Category/मद केटेगरी	endotracheal tubes or reinforced endotracheal tube plain (Q3)
MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है	Yes
Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है	Yes
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	OEM Authorization Certificate, Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid
Primary product category	endotracheal tubes or reinforced endotracheal tube plain
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No

Bid Details/बिड विवरण**Evaluation Method/मूल्यांकन पद्धति**

Item wise evaluation/

EMD Detail/ईएमडी विवरण

Required/आवश्यकता

No

ePBG Detail/ईपीबीजी विवरण

Required/आवश्यकता

No

Splitting/विभाजन

Bid splitting not applied/बोली विभाजन लागू नहीं किया गया.

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता

No

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता

Yes

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
3. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25%(selected by Buyer) percentage of total QUANTITY.
4. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any

impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

Evaluation Method (Item Wise Evaluation Method)

Contract will be awarded schedulewise and the determination of L1 will be done separately for each schedule. The details of item-consignee combination covered under each schedule are as under:

Evaluation Schedules	Item/Category	Quantity
Schedule 1	Endotracheal Tubes Or Reinforced Endotracheal Tube Plain	200
Schedule 2	Endotracheal Tubes Or Reinforced Endotracheal Tube Plain	200
Schedule 3	Endotracheal Tubes Or Reinforced Endotracheal Tube Plain	400
Schedule 4	Endotracheal Tubes Or Reinforced Endotracheal Tube Plain	1000
Schedule 5	Endotracheal Tubes Or Reinforced Endotracheal Tube Plain	300
Schedule 6	Endotracheal Tubes Or Reinforced Endotracheal Tube Plain	1500
Schedule 7	Endotracheal Tubes Or Reinforced Endotracheal Tube Plain	50
Schedule 8	Endotracheal Tubes Or Reinforced Endotracheal Tube Plain	250
Schedule 9	Endotracheal Tubes Or Reinforced Endotracheal Tube Plain	350

Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (200 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	Size (I.D) (mm)	6.0 (millimeter)
	Tube Length (mm)	230
MATERIAL	Material of ET Tube (Non-toxic, Medical Grade)	100% PVC
CONSTRUCTION	Cuffed	Yes
	Cuff Shape	Barrel

Additional Specification Parameters - Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (200 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
External features	Bevel facing left, non traumatic bevel, Murphy's eye, Magill's curvature
External Markings	Black mark for depth placement, radio-opaque line, 1cm marks printed up to 26 cm, printed ID and ED on the tube
Cuff requirements	High volume low pressure, uniform inflation
Pilot balloon	Spring-loaded inflation valve
15 mm connector at end of tube	Yes

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	200	15

Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (200 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* [As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	Size (I.D) (mm)	6.5 (millimeter)
	Tube Length (mm)	60
MATERIAL	Material of ET Tube (Non-toxic, Medical Grade)	100% PVC
CONSTRUCTION	Cuffed	Yes, No
	Cuff Shape	NA

Additional Specification Parameters - Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (200 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Cuffed/ Cuff requirements	Yes/ High-volume low-pressure, uniform inflation
External features	Bevel facing left, non traumatic bevel, Murphy's eye, Magill's curvature
External Markings	Black mark for depth placement, radio-opaque line, 1cm marks printed up to 26 cm, printed ID and ED on the tube
Pilot balloon	Spring-loaded inflation valve
15 mm connector at end of tube	Yes

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	200	15

Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (400 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* [As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	Size (I.D) (mm)	7.0 (millimeter)
	Tube Length (mm)	300
MATERIAL	Material of ET Tube (Non-toxic, Medical Grade)	100% PVC
CONSTRUCTION	Cuffed	Yes
	Cuff Shape	Barrel

Additional Specification Parameters - Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (400 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
External features	Bevel facing left, non traumatic bevel, Murphy's eye, Magill's curvature
External Markings	Black mark for depth placement, radio-opaque line, 1cm marks printed up to 28 cm, printed ID and ED on the tube
Cuff requirements	High-volume low-pressure, uniform inflation
Pilot balloon	Spring-loaded inflation valve
15 mm connector at end of tube	Yes

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	400	15

Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (1000 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* [As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	Size (I.D) (mm)	7.5 (millimeter)
	Tube Length (mm)	330
MATERIAL	Material of ET Tube (Non-toxic, Medical Grade)	100% PVC
CONSTRUCTION	Cuffed	Yes
	Cuff Shape	Barrel

Additional Specification Parameters - Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (1000 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
External features	Bevel facing left, non traumatic bevel, Murphy's eye, Magill's curvature
External Markings	Black mark for depth placement, radio-opaque line, 1cm marks printed up to 28 cm, printed ID and ED on the tube
Cuff requirements	High-volume low-pressure, uniform inflation
Pilot balloon	Spring-loaded inflation valve
15 mm connector at end of tube	Yes

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	1000	15

Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (300 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* [As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	Size (I.D) (mm)	8.0 (millimeter)
	Tube Length (mm)	300
MATERIAL	Material of ET Tube (Non-toxic, Medical Grade)	100% PVC
CONSTRUCTION	Cuffed	Yes
	Cuff Shape	Barrel

Additional Specification Parameters - Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (300 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
External features	Bevel facing left, non traumatic bevel, Murphy's eye, Magill's curvature
External Markings	Black mark for depth placement, radio-opaque line, 1cm marks printed up to 28 cm, printed ID and ED on the tube
Cuff requirements	High-volume low-pressure, uniform inflation
Pilot balloon	Spring-loaded inflation valve
15 mm connector at end of tube	Yes

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	300	15

Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (1500 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* [As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	Size (I.D) (mm)	8.5 (millimeter)
	Tube Length (mm)	330
MATERIAL	Material of ET Tube (Non-toxic, Medical Grade)	100% PVC
CONSTRUCTION	Cuffed	Yes
	Cuff Shape	Barrel

Additional Specification Parameters - Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (1500 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
External features	Bevel facing left, non traumatic bevel, Murphy's eye, Magill's curvature
External Markings	Black mark for depth placement, radio-opaque line, 1cm marks printed up to 30 cm, printed ID and ED on the tube
Cuff requirements	High-volume low-pressure, uniform inflation
Pilot balloon	Spring-loaded inflation valve
15 mm connector at end of tube	Yes

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	1500	15

Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (50 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* [As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	Size (I.D) (mm)	9.0 (millimeter)
	Tube Length (mm)	350
MATERIAL	Material of ET Tube (Non-toxic, Medical Grade)	100% PVC
CONSTRUCTION	Cuffed	Yes
	Cuff Shape	Other

Additional Specification Parameters - Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (50 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
External features	Bevel facing left, non traumatic bevel, Murphy's eye, Magill's curvature
External Markings	Black mark for depth placement, radio-opaque line, 1cm marks printed up to 30 cm, printed ID and ED on the tube
Cuff requirements	High-volume low-pressure, uniform inflation
Pilot balloon	Spring-loaded inflation valve
15 mm connector at end of tube	Yes
External Markings	Black mark for depth placement, radio-opaque line, 1cm marks printed up to 30 cm, printed ID and ED on the tube
Cuff requirements	High-volume low-pressure, uniform inflation
Pilot balloon	Spring-loaded inflation valve
15 mm connector at end of tube	Yes

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	50	15

Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (250 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	Size (I.D) (mm)	5.5 (millimeter)
	Tube Length (mm)	280
MATERIAL	Material of ET Tube (Non-toxic, Medical Grade)	100% PVC
CONSTRUCTION	Cuffed	Yes
	Cuff Shape	Barrel

Additional Specification Parameters - Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (250 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
External features	Bevel facing left, non traumatic bevel, Murphy's eye, Magill's curvature
External Markings	Black mark for depth placement, radio-opaque line, 1cm marks printed up to 24 cm, printed ID and ED on the tube
Cuff requirements	High-volume low-pressure, uniform inflation
Pilot balloon	Spring-loaded inflation valve
15 mm connector at end of tube	Yes

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	250	15

Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (350 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम कैटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	Size (I.D) (mm)	5.0 (millimeter)
	Tube Length (mm)	180
MATERIAL	Material of ET Tube (Non-toxic, Medical Grade)	100% PVC
CONSTRUCTION	Cuffed	Yes
	Cuff Shape	Barrel

Additional Specification Parameters - Endotracheal Tubes Or Reinforced Endotracheal Tube Plain (350 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
External features	Bevel facing left, non traumatic bevel, Murphy's eye, Magill's curvature
External Markings	Black mark for depth placement, radio-opaque line, 1cm marks printed up to 22 cm, printed ID and ED on the tube
Cuff requirements	High-volume low-pressure, uniform inflation
Pilot balloon	Spring-loaded inflation valve
15 mm connector at end of tube	Yes

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	350	15

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें**1. Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

2. Generic

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

3. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

- (1) Sample (1 no each) should be provided for technical evaluation on or before 01.07.2023 at 1 1.00 am.

Please specify bid number on the top of the packet.

Delivery Address:

Sr.Purchase& Stores Officer, AMCHSS, SCTIMST, Medical College P O, Thiruvananthapuram PIN-69 5011

(2) Provide details of GST rate and HSN Code

(3) Bidder should comply all technical specifications mentioned in the bid document. Format of compliance statement is attached in clause-4 of bid document.

(4) All the items should be sterile and have minimum 2 years of shelf life from the date of supply.

4. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

5. **Purchase Preference (Centre)**

Indian suppliers of this item are not allowed to participate and/ or compete in procurement by some foreign governments. Bidders / products from such countries are not eligible / not allowed to participate in this bid in terms of clause 1 (d) of Public Procurement (Preference to Make in India) Order, 2017

6. **Generic**

Manufacturer Authorization:Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

7. **Generic**

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

8. **Generic**

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. Any clause(s) incorporated by the Buyer regarding following shall be treated as null and void and would not be considered as part of bid:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process.
9. Mandating foreign / international certifications even in case of existence of Indian Standards without

specifying equivalent Indian Certification / standards.

10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---