



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2023/B/3363726
Dated/दिनांक : 15-05-2023

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	25-05-2023 14:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	25-05-2023 14:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Science And Technology
Department Name/विभाग का नाम	Department Of Science And Technology (dst)
Organisation Name/संगठन का नाम	Sree Chitra Tirunal Institute For Medical Sciences And Technology (sctimst)
Office Name/कार्यालय का नाम	Thiruvananthapuram
Total Quantity/कुल मात्रा	210
Item Category/मद केटेगरी	LARYNGEAL MASK AIRWAYS (Q3)
MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है	Yes
Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है	Yes
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	OEM Authorization Certificate, Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Single Packet Bid
Primary product category	LARYNGEAL MASK AIRWAYS
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No

Bid Details/बिड विवरण**Evaluation Method/मूल्यांकन पद्धति**

Total value wise evaluation

EMD Detail/ईएमडी विवरण

Required/आवश्यकता

No

ePBG Detail/ईपीबीजी विवरण

Required/आवश्यकता

No

Splitting/विभाजन

Bid splitting not applied/बोली विभाजन लागू नहीं किया गया.

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता

No

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता

Yes

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
3. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25%(selected by Buyer) percentage of total QUANTITY.

LARYNGEAL MASK AIRWAYS (30 pieces)

Brand Type/ब्रांड का प्रकार	Registered Brand	
Technical Specifications/तकनीकी विशिष्टियाँ		
<u>* As per GeM Category Specification/जेम कैटेगरी विशिष्टि के अनुसार</u>		
Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
MATERIAL	Material of Mask (Non-toxic , Medical Grade Latex free) , soft and smooth	Yes
	Material of Tube (Non-toxic , medical grade)	100% Silicon
GENERAL	Laryngeal Mask Type	Single use airway mask
	Usage	Pediatric
	Laryngeal Mask Airway Size(Number)	1
	Integraged bite block and double cuff	Yes
	Method of Sterilization	Others
CONSTRUCTION	Separate Inflation line to enable the use to move away the line from the teeth of the patient to prevent accidental cuff deflation	Yes
	Shall have a port for respiration and port size should be adequate enough for Ryle's tube insertion	Yes
	With a special curve moulded directly into the airway tube with a unique easy glide surface	Yes
	Special reinforced tip which helps to prevent folds during insertion and plugs the upper esophageal sphincter	No
	Cuff and airway tube moulded as a single unit	No
DIMENSION	Length of Airway tube(mm)	291
	Inner Diameter of Tube(mm)	6.2
MARKING AND PACKING	Type of Packing	Single Individually packed in soft wrap

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	30	15

LARYNGEAL MASK AIRWAYS (30 pieces)

Brand Type/ब्रांड का प्रकार	Registered Brand
-----------------------------	------------------

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
MATERIAL	Material of Mask (Non-toxic , Medical Grade Latex free) , soft and smooth	Yes
	Material of Tube (Non-toxic , medical grade)	100% PVC
GENERAL	Laryngeal Mask Type	Single use airway mask
	Usage	Pediatric
	Laryngeal Mask Airway Size(Number)	1.5
	Integraged bite block and double cuff	Yes
	Method of Sterilization	Others
CONSTRUCTION	Separate Inflation line to enable the use to move away the line from the teeth of the patient to prevent accidental cuff deflation	Yes
	Shall have a port for respiration and port size should be adequate enough for Ryle's tube insertion	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	With a special curve moulded directly into the airway tube with a unique easy glide surface	Yes
	Special reinforced tip which helps to prevent folds during insertion and plugs the upper esophageal sphincter	No
	Cuff and airway tube moulded as a single unit	No
DIMENSION	Length of Airway tube(mm)	291
	Inner Diameter of Tube(mm)	8
MARKING AND PACKING	Type of Packing	Single Individually packed in soft wrap

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	30	15

LARYNGEAL MASK AIRWAYS (50 pieces)

Brand Type/ब्रांड का प्रकार	Registered Brand
-----------------------------	------------------

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
---------------	------------------------------------	--

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
MATERIAL	Material of Mask (Non-toxic , Medical Grade Latex free) , soft and smooth	Yes
	Material of Tube (Non-toxic , medical grade)	100% Silicon
GENERAL	Laryngeal Mask Type	Single use airway mask
	Usage	Pediatric
	Laryngeal Mask Airway Size(Number)	2
	Integraged bite block and double cuff	Yes
	Method of Sterilization	Gamma
CONSTRUCTION	Separate Inflation line to enable the use to move away the line from the teeth of the patient to prevent accidental cuff deflation	Yes
	Shall have a port for respiration and port size should be adequate enough for Ryle's tube insertion	Yes
	With a special curve moulded directly into the airway tube with a unique easy glide surface	Yes
	Special reinforced tip which helps to prevent folds during insertion and plugs the upper esophageal sphincter	No
	Cuff and airway tube moulded as a single unit	Yes
DIMENSION	Length of Airway tube(mm)	189
	Inner Diameter of Tube(mm)	7.8
MARKING AND PACKING	Type of Packing	Single Individually packed in soft wrap

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	50	15

LARYNGEAL MASK AIRWAYS (30 pieces)

Brand Type/ब्रांड का प्रकार	Registered Brand
-----------------------------	------------------

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
MATERIAL	Material of Mask (Non-toxic , Medical Grade Latex free) , soft and smooth	Yes
	Material of Tube (Non-toxic , medical grade)	100% Silicon
GENERAL	Laryngeal Mask Type	Single use airway mask
	Usage	Pediatric
	Laryngeal Mask Airway Size(Number)	2.5
	Integraged bite block and double cuff	Yes
	Method of Sterilization	Others
CONSTRUCTION	Separate Inflation line to enable the use to move away the line from the teeth of the patient to prevent accidental cuff deflation	Yes
	Shall have a port for respiration and port size should be adequate enough for Ryle's tube insertion	Yes
	With a special curve moulded directly into the airway tube with a unique easy glide surface	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Special reinforced tip which helps to prevent folds during insertion and plugs the upper esophageal sphincter	No
	Cuff and airway tube moulded as a single unit	Yes
DIMENSION	Length of Airway tube(mm)	291
	Inner Diameter of Tube(mm)	8
MARKING AND PACKING	Type of Packing	Single Individually packed in blister pack

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	30	15

LARYNGEAL MASK AIRWAYS (30 pieces)

Brand Type/ब्रांड का प्रकार	Registered Brand
-----------------------------	------------------

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
MATERIAL	Material of Mask (Non-toxic , Medical Grade Latex free) , soft and smooth	Yes
	Material of Tube (Non-toxic , medical grade)	100% Silicon

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GENERAL	Laryngeal Mask Type	Single use airway mask
	Usage	Adult
	Laryngeal Mask Airway Size(Number)	3
	Integraged bite block and double cuff	Yes
	Method of Sterilization	ETO
CONSTRUCTION	Separate Inflation line to enable the use to move away the line from the teeth of the patient to prevent accidental cuff deflation	Yes
	Shall have a port for respiration and port size should be adequate enough for Ryle's tube insertion	Yes
	With a special curve moulded directly into the airway tube with a unique easy glide surface	Yes
	Special reinforced tip which helps to prevent folds during insertion and plugs the upper esophageal sphincter	Yes
	Cuff and airway tube moulded as a single unit	Yes
DIMENSION	Length of Airway tube(mm)	291
	Inner Diameter of Tube(mm)	10.2
MARKING AND PACKING	Type of Packing	Single Individually packed in blister pack

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	30	15

LARYNGEAL MASK AIRWAYS (20 pieces)

Brand Type/ब्रांड का प्रकार	Registered Brand
-----------------------------	------------------

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
MATERIAL	Material of Mask (Non-toxic , Medical Grade Latex free) , soft and smooth	Yes
	Material of Tube (Non-toxic , medical grade)	100% Silicon
GENERAL	Laryngeal Mask Type	Single use airway mask
	Usage	Adult
	Laryngeal Mask Airway Size(Number)	4
	Integraged bite block and double cuff	Yes
	Method of Sterilization	ETO
CONSTRUCTION	Separate Inflation line to enable the use to move away the line from the teeth of the patient to prevent accidental cuff deflation	Yes
	Shall have a port for respiration and port size should be adequate enough for Ryle's tube insertion	Yes
	With a special curve moulded directly into the airway tube with a unique easy glide surface	Yes
	Special reinforced tip which helps to prevent folds during insertion and plugs the upper esophageal sphincter	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Cuff and airway tube moulded as a single unit	Yes
DIMENSION	Length of Airway tube(mm)	292
	Inner Diameter of Tube(mm)	10.2
MARKING AND PACKING	Type of Packing	Single Individually packed in blister pack

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	20	15

LARYNGEAL MASK AIRWAYS (20 pieces)

Brand Type/ब्रांड का प्रकार	Registered Brand
-----------------------------	------------------

Technical Specifications/तकनीकी विशिष्टियाँ

[* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
MATERIAL	Material of Mask (Non-toxic , Medical Grade Latex free) , soft and smooth	Yes
	Material of Tube (Non-toxic , medical grade)	100% PVC
GENERAL	Laryngeal Mask Type	Single use airway mask
	Usage	Adult
	Laryngeal Mask Airway Size(Number)	5

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Integraged bite block and double cuff	Yes
	Method of Sterilization	ETO
CONSTRUCTION	Separate Inflation line to enable the use to move away the line from the teeth of the patient to prevent accidental cuff deflation	Yes
	Shall have a port for respiration and port size should be adequate enough for Ryle's tube insertion	Yes
	With a special curve moulded directly into the airway tube with a unique easy glide surface	Yes
	Special reinforced tip which helps to prevent folds during insertion and plugs the upper esophageal sphincter	Yes
	Cuff and airway tube moulded as a single unit	Yes
DIMENSION	Length of Airway tube(mm)	Any applicable numeric value
	Inner Diameter of Tube(mm)	Any applicable numeric value
MARKING AND PACKING	Type of Packing	Single Individually packed in soft wrap

Additional Specification Parameters - LARYNGEAL MASK AIRWAYS (20 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Length of Airway Tube(mm)	190
Inner Diameter of Tube(mm)	10

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषितो/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shiju V S	695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum	20	15

Special terms and conditions-Version:1 effective from 21-10-2022 for category LARYNGEAL MASK AIRWAYS

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules made there under as amended till date will always be applicable. This will include all notifications issued by *Central Drugs Standard Control Organisation (CDSCO)*, Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. The price offered by the seller shall not, in any case exceed the DPCO controlled price, if any, fixed by the Central/State Government, the Maximum Retail Price (MRP) and the selling price. The seller must reduce the prices if there is any reduction in DPCO ceiling price, if any.
4. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e., ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

2. Generic

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

3. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

- (1) Sample should be provided for technical evaluation on or before 26.05.2023 at 11.00 am.

Please specify bid number on the top of the packet.

Delivery Address:

Sr.Purchase& Stores Officer, AMCHSS, SCTIMST, Medical College P O, Thiruvananthapuram PIN-69 5011

- (2) Provide details of GST rate and HSN Code

4. **Purchase Preference (Centre)**

Indian suppliers of this item are not allowed to participate and/ or compete in procurement by some foreign governments. Bidders / products from such countries are not eligible / not allowed to participate in this bid in terms of clause 1 (d) of Public Procurement (Preference to Make in India) Order, 2017

5. **Generic**

Manufacturer Authorization: Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

6. **Generic**

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

7. **Generic**

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. Any clause(s) incorporated by the Buyer regarding following shall be treated as null and void and would not be considered as part of bid:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process.
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---